

# DROP TEST CHLORIDE (1 drop = 10 ppm)

Instr. #5067O

## COMPONENTS:

- |               |   |
|---------------|---|
| 1 x 5067O     | Instruction   |
| 1 x 9198O     | Sample Tube, Graduated (25 mL) w/ cap & orange dot, plastic |
| 1 x R-0630-C  | Chromate Indicator, 2 oz, DB                                |
| 1 x R-0638O-A | Phenolphthalein Indicator, .75 oz w/ orange cap, DB         |
| 1 x R-0687O-C | Sulfuric Acid .12N, 2 oz w/ orange cap, DB                  |
| 1 x R-0706-C  | Silver Nitrate Reagent, 2 oz, DB                            |

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE  
800-TEST KIT (800-837-8548).

## PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.  
KEEP REAGENTS AWAY FROM CHILDREN.

NOTE: When dispensing reagents from dropper bottles, **always** hold bottle in a vertical position.

### Chloride Test

NOTE: When sulfite content of sample water to be tested exceeds 10 ppm, the sulfite should be oxidized to prevent interference in test. A 25 mL water sample is first adjusted to the appropriate pH, then 1 mL (or 25 drops) of R-0649 Hydrogen Peroxide Solution (sold separately) is added and thoroughly mixed. Continue with the rest of the procedure.

1. Rinse and fill 25 mL sample tube (#9198O) to 25 mL mark with water to be tested.
2. Add 2 drops R-0638O Phenolphthalein Indicator. Swirl to mix. If sample turns pink (Fig. 1), add R-0687O Sulfuric Acid .12N dropwise, swirling after each drop, until color changes from pink to colorless.

3. Add 5 drops R-0630 Chromate Indicator. Swirl to mix. Sample will turn yellow (Fig. 2).

4. Add R-0706 Silver Nitrate Reagent dropwise, swirling and counting after each drop, until color changes from yellow to milky salmon (brick red) (Fig. 3).

NOTE: A white precipitate will form as R-0706 Silver Nitrate Reagent is added to the sample. Do not add enough R-0706 Silver Nitrate Reagent to give a brown color. First change from yellow to a milky salmon (brick red) is the endpoint.

5. Multiply drops of R-0706 Silver Nitrate Reagent by 10. Record as parts per million (ppm) chloride (Cl<sup>-</sup>).



Fig. 1



Fig. 2



Fig. 3



31 Loveton Circle, Sparks, MD 21152 USA  
800-TEST KIT (837-8548) • 410-472-4340

082517