

COLOR COMPARISON TEST TOTAL CHLORINE

Instr. #5509

MIDGET COMPARATORS*: RANGE:

| | |
|------|---------------------------------|
| 9031 | Chlorine 0.2-0.8 ppm/pH 6.8-8.0 |
| 9039 | Chlorine 0-1.0 ppm |
| 9042 | Chlorine 2-8.0 ppm/pH 7.2-8.4 |
| 9045 | Chlorine 0.5-2.0 ppm/pH 6.8-8.0 |
| 9047 | Chlorine 0-2.0 ppm |

*Kit includes only one comparator.

REAGENT:

R-0600-A Orthotolidine*, .75 oz

APPARATUS:

| | |
|----------|---|
| 1 x 3243 | Cap, Test Cell (11.5 mL), plastic (or #3267, 5 mL) |
| 1 x 4024 | Test Cell, Calibrated (11.5 mL), plastic (or #4025, 5 mL) |
| 1 x 4028 | Pipet, Calibrated (0.5 mL), w/ cap, plastic |
| 1 x 5509 | Instruction |
| 1 x 6002 | Brush, Test Cell |

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE
800-TEST KIT (800-837-8548).

PROCEDURE:

**CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.**

1. Rinse and fill test cell (#4024 or #4025) to 11.5 mL (or 5 mL) mark with water to be tested.

NOTE: Use test cell #4025 for high chlorine (2-8 ppm).

2. Using 0.5 mL pipet (#4028), add 0.5 mL R-0600 Orthotolidine. Cap and mix.

3. Wipe dry and place in comparator WITH FROSTED SIDE FACING OPERATOR.

4. Match color in test cell with a color standard. Record as parts per million (ppm) total chlorine (Cl_2).

*WARNING: Orthotolidine (R-0600) contains 5-10% w/w hydrochloric acid, a strong corrosive acid, and 0.1-1% w/w orthotolidine, a suspected carcinogen.



31 Loveton Circle, Sparks, MD 21152 USA
800-TEST KIT (837-8548) • 410-472-4340

011018